

 Eskom	SCOPE OF WORK	Grootvlei Power Station
---	----------------------	-------------------------

To supply, deliver, Ion Chromatography instrument then train staff on how to use it at Grootvlei power station water laboratory. Service instrument for period of 60 months on as and when required basis.

Document Identifier: **GVL/0715**

Alternative Reference Number: **N/A**

Area of Applicability: **Grootvlei Power Station**

Functional Area: **Chemical Services**

Revision: **1**

Total Pages: **7**

Next Review Date: **AS REQUIRED**

Disclosure Classification: **Controlled Disclosure**

Compiled by

Supported by

Authorized by

Date:

Date:

Date:

CONTENTS

Page

1.	INTRODUCTION	3
2.	SUPPORTING CLAUSES	3
2.2	Normative / Informative References	4
2.2.1	Normative	4
2.2.2	Informative	4
2.3	Definitions	4
2.4	Roles and Responsibilities	4
2.5	Process for Monitoring	5
2.6	Related/Supporting Documents	5
3.	EMPLOYERS OBJECTIVES AND PURPOSE OF THE WORKS.	5
3.1	SCOPE OF WORK Price List	5
4.	AUTHORISATION	6
5.	REVISIONS	7
6.	DEVELOPMENT TEAM	7
7.	ACKNOWLEDGEMENTS	7

CONTROLLED DISCLOSURE

I. INTRODUCTION

Grootvlei Power Station consists of 3 wet cooled units which produce 200MW each at full load, during the process of electricity generation, the station uses anion results from Ion Chromatography to minimize corrosive anions. These ions contribute to turbine corrosion, boiler tube leaks, cooling tower structures corrosion and permanent scaling in condenser. Grootvlei power station require this instrument to be able to perform the necessary chemistry monitoring.

ISO 9001 quality management system (QMS) states that all the instruments used for testing to be serviced and calibrated for reliable results. Grootvlei Power Station, Chemical Services will require the supplied instrument to be serviced for period of 60 months on as and when required basis to comply with ISO 9001 QMS.

2. SUPPORTING CLAUSES

2.1 Scope

The scope of work is outline below:

Grootvlei power station require Ion Chromatography instrument for analysis of water, to be supplied and installed then training of staff. The instrument requires service and maintenance as per ISO 9001 Quality management on a yearly basis and as and when service is required during breakdowns. The supplier must supply us with consumables (as per price list) to run the instrument for a period of 60 months and must be available during callouts as per clause X17 appendix A.

Supply instrument as per below specification

- a) Instrument must be able to analyse water samples from parts per billion (PPB) to parts per million (PPM). Must be able to produce repeatable results at 2.5ppb and less.
- b) Instrument must use software and computer to run.
- c) There must be an automated sampler with sample rack and operating with no risk of contamination between low level and high level. Strictly high level automated.
- d) Must be interchangeable between low level and high level or Separate high level and low-level systems.
- e) It must be a bench top analyser.
- f) Must be able to accept a wide range calibration for low level and for high level.
- g) Must be able to detect conductivity.
- h) Must use analytical columns and guard columns.

Scope of Service.

Service of Ion Chromatography.

- ☐ Ensure flow of reagents is happening as per design.
- ☐ Ensure the sample analysis is occurring according to design.
- ☐ Ensure instrument is producing required results by running standards.
- ☐ Ensure all the deviations are corrected.

CONTROLLED DISCLOSURE

- ☐ Change spares that are due for replacements during service.
- ☐ Supply spares for the instrument.
- ☐ Ensure instrument is running optimally.
- ☐ Provide calibration certificates.
- ☐ Provide calibration and service report detailing all work done.
- ☐ Provide sticker on the side of the instrument indicating date serviced, next service date and signature of service engineer.
- ☐ Perform detailed service of the instrument to check all the components that should be checked on annual service and peaks are integrating as per requirement.

Applicability

This document shall apply to Eskom Grootvlei Powe Station, Chemical Services, Water Laboratory.

2.2 Normative / Informative References

Parties using this document shall apply the most recent edition of the documents listed in the following paragraphs.

2.2.1 Normative

- [1] ISO 9001: 2015 – Quality Management Systems requirements.

2.2.2 Informative

- [1] Not applicable

2.3 Definitions

- [1] Not applicable

Abbreviation	Description
QMS	Quality Management System
IC	Ion Chromatography
PPB	Parts Per Billion
PPM	Parts Per Million

2.4 Roles and Responsibilities

It is the responsibility of Chemical Services supervisor to ensure the instrument supplied is of correct specification, serviced and calibrated.

CONTROLLED DISCLOSURE

2.5 Process for Monitoring

- The process for monitoring will be governed by calibration and service schedules.

2.6 Related/Supporting Documents

- Calibration certificates

3. EMPLOYERS OBJECTIVES AND PURPOSE OF THE WORKS.

To ensure the instrument is serviced, calibrated and is reliable to conduct test as per ISO 9001 QMS.

3.1 SCOPE OF WORK Price List

Item nr	Description	Unit	Expected Quantity	Rate	Price
1	Supply, Deliver and Install Benchtop Ion Chromatography Instrument.				
1.1	Supply of new Ion Chromatography instrument (Including delivery, installation, and Training staff), Package includes: <ul style="list-style-type: none"> Computer Instrument(s) able to run high- and low-level water samples. Windows-based operating software. External PC Tower, keyboard, and mouse, 19". Operating supplies. 230VAC, 50/60Hz, 30Amps.	EA	1		
Total					R
2	Service of Ion Chromatography Instrument and Supply Consumables				
1	Service of Ion Chromatography instrument	Year	5		
2	Service labour	Hr	80		

CONTROLLED DISCLOSURE

To supply, deliver, Ion Chromatography instrument then train staff on how to use it at Grootvlei power station water laboratory. Service instrument for period of 60 months on as and when required basis.

Unique Identifier: **GVL/0715**
Revision: **1**
Page: **6 of 7**

3	Travelling cost	Km	1250		
4	SHE file	Sum	1		
5	Medicals	EA	10		
6	PPE (Safety Shoe, Goggles, and lab coat)	EA	10		
7	Police clearance	EA	10		
8	Overnight accommodation	EA	5		
Total					R
3	Consumables to run Ion Chromatography instrument for 5 years. We run 25 samples per day. (12 low level and 13 high level samples per day). Add the list of consumables that we will be able to use on a day to day running of instrument for five years.				
1		Ea			
2		Ea			
3		Ea			
4		Ea			
5		Ea			
6		Ea			
7		Ea			
8		Ea			
9		Ea			
10		Ea			
11		Ea			
12		Ea			
13		Ea			
n					

4. AUTHORISATION

This document has been seen and accepted by:

CONTROLLED DISCLOSURE

When downloaded from the document management system, this document is uncontrolled and the responsibility rests with the user to ensure it is in line with the authorized version on the system. No part of this document may be reproduced in any manner or form by third parties without the written consent of Eskom Holdings SOC Ltd, © copyright Eskom Holdings SOC Ltd, Reg No 2002/015527/30

To supply, deliver, Ion Chromatography instrument then train staff on how to use it at Grootvlei power station water laboratory. Service instrument for period of 60 months on as and when required basis.

Unique Identifier: **GVL/0715**

Revision: **1**

Page: **7 of 7**

5. REVISIONS

Date	Revision	Compiler	Remarks
October 2025	01	Mpho Netshidzati	New document

6. DEVELOPMENT TEAM

The following people were involved in the development of this document:

- Mpho Netshidzati

7. ACKNOWLEDGEMENTS

Not applicable.

CONTROLLED DISCLOSURE

When downloaded from the document management system, this document is uncontrolled and the responsibility rests with the user to ensure it is in line with the authorized version on the system. No part of this document may be reproduced in any manner or form by third parties without the written consent of Eskom Holdings SOC Ltd, © copyright Eskom Holdings SOC Ltd, Reg No 2002/015527/30